



[WWW.SECURITYSYSTEMS.NO](http://WWW.SECURITYSYSTEMS.NO)

## MARCA BLANCA BLACKBODY-TH



Blackbody - Calibration device for thermographic cameras - Infrared emission of 35°C ~ 60°C - Stability  $\pm 0.1\sim 0.2^{\circ}\text{C}/\text{h}$  - Emissivity  $0.97 \pm 0.02$  - Guarantees a measurement accuracy of  $\pm 0.3^{\circ}\text{C}$

- Blackbody
- Guarantees a measurement accuracy of  $\pm 0.3^{\circ}\text{C}$
- Temperature range: 35°C ~ 60°C
- Accuracy  $\pm 0.2^{\circ}\text{C}$
- Radiation area 78x78 mm
- Stability  $\pm 0.1\sim 0.2^{\circ}\text{C}/\text{h}$
- Emissivity  $0.97 \pm 0.02$